

## EU DECLARATION OF CONFORMITY

We, Leica Biosystems Nussloch GmbH, Heidelberger Str. 17-19, 69226 Nussloch, Germany

declare on our own responsibility, that the device

## HistoCore BIOCUT R

complies with ANNEX I essential health and safety requirements relating to the design and construction of machinery

Directive 2006/42/EC of the European Parliament and of the Council of 17 May 2006 on machinery, and amending Directive 95/16/EC (recast) (OJ L 157, 9.6.2006, p. 24-86)

EN ISO 12100:2010 EN ISO 14971:2012 EN 61010-1:2010

Quality Management System: Certified according to EN ISO 9001:2015

Manufacturing site:

Leica Microsystems Ltd. Shanghai, Building 1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade, 201206 Shanghai, People's

Republic of China

Nussloch, 22.06.2020

Andreas Hahr

Senior Director Core Histology

Robert Gropp RA/QA Manager