



Certificate of Compliance

Certificate: 80145472

Master Contract: 217333

Project: 80155758

Date Issued: 2023-02-27

Issued To: Leica Biosystems Nussloch GmbH
Heidelbergerstrasse 17-19
Nussloch., Baden-Württemberg, 69226
Germany

Attention: Titan Tian

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: Tracy Geng
Tracy Geng



PRODUCTS

CLASS - C872106 - ELECTRICAL LABORATORY EQUIPMENT Electrical Laboratory Equipment

CLASS - C872186 - ELECTRICAL LABORATORY EQUIPMENT Certified to US Standards

Water Bath, model HistoCore Water Bath M, rated 100-120V~, 8A Max. or 220-240V~, 5A Max. 50/60 Hz,

Notes:

1. The above model is cord connected, Equipment Class I, Pollution Degree 2, Overvoltage category II
2. Mode of operation: Continuous
3. Environmental Conditions: Normal: 15 to 40°C, 2000 m max, 20% to 80% non-condensing as specified by manufacturer



Certificate: 80145472
Project: 80155758

Master Contract: 217333
Date Issued: 2023-02-27

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA C22.2 No. 61010-1-12, UPD1: 2015,
UPD2: 2016, AMD1: 2018

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CSA C22.2 No. 61010-2-010:2019

- Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-010: Particular requirements for laboratory equipment for the heating of materials

CSA-C22.2 No. 61010-2-101:19

- Safety requirements for electrical equipment for measurement, control and laboratory use — Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment

UL Standards:

UL 61010-1 3rd edition (2012), AMD1: 2018 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL 61010-2-010 4th edition (2019)

- Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-010: Particular requirements for laboratory equipment for the heating of materials

UL 61010-2-101 Ed.3 (2019)

- Safety requirements for electrical equipment for measurement, control and laboratory use — Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment

Notes:

Products certified under Class C872106, C872186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80145472

Master Contract: 217333

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80155758	2023-02-27	Update report 80145472 to change model name from HistoCore Water Bath to HistoCore Water Bath M and add standards CSA-C22.2 No. 61010-2-101:19 and UL 61010-2-101 Ed.3 (2019) based on CB acceptance.
80145472	2022-10-24	Initial cCSAus Certification of Water Bath, model HistoCore Water Bath based on CB acceptance.